

Info Sheet INFO 072.E-ed.H

Date of issue: 01/12/19 Replaces version from: 01/03/17

Page 1/1

REPLACING A BONDED TITANIUM FEMALE

- 1. Identify the attachment as an **M2** or **M3** attachment (see INFO 062 and INFO 070).
- 2. Use a sharp (new) cylindrical burr to cut out the female.
- 3. Cut in circular movements, holding the burr parallel to the female wall.
- 4. When the metal is cut through, the **CEKA SITE** material appears. Now also remove the remaining **CEKA SITE** with a cylindrical diamond burr.
- 5. Do not damage the upper edge of the cast part of the post-coping.
- 6. The cavity in the post-coping must remain cylindrical.
- 7. With **M2** or **M3** attachments, the following parts are assembled:
 - M2: RE H 14 (accessory) + RE 0031 (spring pin) + RE 0096 (space maintainer)
 - + RE 4500 TI (titanium female);
 - M3: H 14 (accessory) + 694 C (spring pin) + 694 B (space maintainer) + 694 TI (titanium female).
- 8. Check the assembled parts to be sure the female fits fully into the cavity and make any necessary corrections.
- 9. Clean the cavity and the outside of the titanium female.
- Apply a small amount of CEKA SITE in the cavity and press the female with the assembled parts into the cavity (see INFO 074).
- 11. Do not move these parts during the entire setting time (10 minutes).
- 12. Remove the accessories with the spring pin and the excess **CEKA SITE** bonding composite.

Side effects, warnings and precautions

- The attachments are intended for single use.
- The products are non-sterile.
- There is a risk of poor fit when patient conditions change.
- Bacterial adhesion can be avoided by applying hygiene measures.
- Inappropriate use or bad manufacturing can lead to premature wear of the attachments.
- The functionality of the attachments will be adversely affected by traumas such as grinding and bruxism.
- For the purpose of traceability we advise you to record the lot number of the applied products in the patient file.
- Do not heat items containing titanium.
- Do not use items containing nickel in case of nickel allergy.
- The accessories RE H 79 and H 35 must be used outside the mouth.